


# MIC-SERIES

compatible to:

*MIS*®  
*C1*®



*A: over surface*

*B: over edge*

## **PLATFORMS:**



*3,3 mm NP*



*3,75 / 4,2 mm SP*



*5,0 mm WP*

# MIC-SERIES

## TITANIUM BASE

incl. screw

MIC 80330000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

MIC 9.S3D3.300



## NT-MILL LAB ANALOG

MIC 5.La3.300




## IMPLANT PICK-UP

for open-tray with integrated screw

MIC TR-00013.3



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

MIC 603300



## DIM ANALOG®

MIC 5.DIM.330



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A10



**NT-PREFORM®**

incl. coloured milling screw

MIC 10.Mi3.300.05

**SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning

MIC 9.S3D3.300

**NT-MILL LAB ANALOG**


for NT-Preform®

MIC 5.La3.300

**COLOURED MILLING SCREW**

for NT-Preform®

MIC 6.Sc3.300

**ABUTMENT SCREW**

Overview of all screwdrivers on pages 344 and 345

MIC 603300

**DIM ANALOG®**

MIC 5.DIM.330

**SLEEVE** (belongs to DIM ANALOG®)  
plaster model

DIM 5.XSL.A10

**ABUTMENT GRIP**

for NT-Holder fixture

W 12.100.000

**CUSTOMIZED ABUTMENT**

milled by NT-iDent

Not available in the USA

TZ15.TiO.Si1



# MIC-SERIES

## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm MIC-2CON 8033001S

GH 2,5 mm MIC-2CON 8033002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm MIC-2CON 8033001M

GH 2,5 mm MIC-2CON 8033002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



## 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



## 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800



**2-CONNECT®-ABUTMENT SET (ANGLED)**


incl. 2-CONnect®-Coping, mounting aid,  
Coping-screw and Male Part Screw, angled 20°,  
Gingiva Height 2,5 mm

A: over surface

MIC-2CON 80330C3S

B: over edge

MIC-2CON 80330A3S

**2-CONNECT®-ABUTMENT**

angled 20°, Gingiva Height 2,5 mm  
Male Part incl. screw,  
mounting aid

A: over surface

MIC-2CON 80330C3M

B: over edge

MIC-2CON 80330A3M

**2-CONNECT®-SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning  
KS 62 = 1,6 mm

2CON 9.S3D1.600

**2-CONNECT®-COPING**

Female Part

2CON F-U

**COPING-SCREW**

Recommended tightening torques: 15 Ncm

KS 62

**MALE PART SCREW**

for angled 20° 2-CONnect® Abutment  
Recommended tightening torques: 20 Ncm

MIC 6.Ms3.300

**2-CONNECT®-DIM ANALOG®**

KS 62 = 1,6 mm

2CON 5.DIM.160



**SLEEVE** (belongs to DIM ANALOG®)  
plaster model

DIM 5.XSL.A30

**2-CONNECT®-IMPLANT PICK-UP**

integrated screw  
KS 62 = 1,6 mm

2CON TR-00031.6

**2-CONNECT®-LAB ANALOG**

KS 62 = 1,6 mm

2CON 5.La1.600



# MIC-SERIES

## TITANIUM BASE

incl. screw

MIC 80374200



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

MIC 9.S3D3.742



## NT-MILL LAB ANALOG

MIC 5.La3.742




## IMPLANT PICK-UP

for open-tray with integrated screw

MIC TR-00023.7



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

MIC 603750



## DIM ANALOG®

MIC 5.DIM.374



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A20



**NT-PREFORM®**

incl. coloured milling screw

MIC 10.Mi3.742.05

**SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning

MIC 9.S3D3.742

**NT-MILL LAB ANALOG**


for NT-Preform®

MIC 5.La3.742

**COLOURED MILLING SCREW**

for NT-Preform®

MIC 6.Sc3.750

**ABUTMENT SCREW**

Overview of all screwdrivers on pages 344 and 345

MIC 603750

**DIM ANALOG®**

MIC 5.DIM.374

**SLEEVE** (belongs to DIM ANALOG®)  
plaster model

DIM 5.XSL.A20

**ABUTMENT GRIP**

for NT-Holder fixture

W 12.200.000

**CUSTOMIZED ABUTMENT**

milled by NT-iDent

Not available in the USA

TZ15.TiO.Si1





# MIC-SERIES

## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm

MIC-2CON 8037421S

GH 2,5 mm

MIC-2CON 8037422S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm

MIC-2CON 8037421M

GH 2,5 mm

MIC-2CON 8037422M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



## 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



## 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800



**2-CONNECT®-ABUTMENT SET (ANGLED)**

incl. 2-CONnect®-Coping, mounting aid,  
Coping-screw and Male Part Screw, angled 20°,  
Gingiva Height 2,5 mm

MIC-2CON 80374A3S

**2-CONNECT®-ABUTMENT**

angled 20°, Gingiva Height 2,5 mm  
Male Part incl. screw,  
mounting aid

MIC-2CON 80374A3M

**2-CONNECT®-SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning  
KS 62 = 1,6 mm

2CON 9.S3D1.600

**2-CONNECT®-COPING**

Female Part

2CON F-U

**COPING-SCREW**

Recommended tightening torques: 15 Ncm

KS 62

**MALE PART SCREW**

for angled 20° 2-CONnect® Abutment  
Recommended tightening torques: 20 Ncm

MIC 6.Ms3.750

**2-CONNECT®-DIM ANALOG®**

KS 62 = 1,6 mm

2CON 5.DIM.160



**SLEEVE** (belongs to DIM ANALOG®)  
plaster model

DIM 5.XSL.A30

**2-CONNECT®-IMPLANT PICK-UP**

integrated screw  
KS 62 = 1,6 mm

2CON TR-00031.6

**2-CONNECT®-LAB ANALOG**

KS 62 = 1,6 mm

2CON 5.La1.600



# MIC-SERIES

## TITANIUM BASE

incl. screw

MIC 80500000



## SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

MIC 9.S3D5.000



## NT-MILL LAB ANALOG

MIC 5.La5.000




## IMPLANT PICK-UP

for open-tray with integrated screw

MIC TR-00035.0



## ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

MIC 603750



## DIM ANALOG®

MIC 5.DIM.500



## SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



**NT-PREFORM®**

incl. coloured milling screw

MIC 10.Mi5.000.05

**SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning

MIC 9.S3D5.000

**NT-MILL LAB ANALOG**


for NT-Preform®

MIC 5.La5.000

**COLOURED MILLING SCREW**

for NT-Preform®

MIC 6.Sc3.750

**ABUTMENT SCREW**

Overview of all screwdrivers on pages 344 and 345

MIC 603750

**DIM ANALOG®**

MIC 5.DIM.500

**SLEEVE** (belongs to DIM ANALOG®)  
plaster model

DIM 5.XSL.A30

**ABUTMENT GRIP**

for NT-Holder fixture

W 12.300.000

**CUSTOMIZED ABUTMENT**

milled by NT-iDent

Not available in the USA

TZ15.TiO.Si1



# MIC-SERIES

## 2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm

MIC-2CON 8050001S

GH 2,5 mm

MIC-2CON 8050002S



## 2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm

MIC-2CON 8050001M

GH 2,5 mm

MIC-2CON 8050002M



## 2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



## 2-CONNECT®-COPING

Female Part

2CON F-U



## COPING-SCREW

Recommended tightening torques: 25 Ncm

KS 61



## 2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



## SLEEVE (belongs to DIM ANALOG®)

plaster model

DIM 5.XSL.A30



## 2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



## 2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm

2CON 5.La1.800



**2-CONNECT®-ABUTMENT SET (ANGLED)**

incl. 2-CONnect®-Coping, mounting aid,  
Coping-screw and Male Part Screw, angled 20°,  
Gingiva Height 2,5 mm

MIC-2CON 80500A3S

**2-CONNECT®-ABUTMENT**

angled 20°, Gingiva Height 2,5 mm  
Male Part incl. screw,  
mounting aid

MIC-2CON 80500A3M

**2-CONNECT®-SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning  
KS 62 = 1,6 mm

2CON 9.S3D1.600

**2-CONNECT®-COPING**

Female Part

2CON F-U

**COPING-SCREW**

Recommended tightening torques: 15 Ncm

KS 62

**MALE PART SCREW**

for angled 20° 2-CONnect® Abutment  
Recommended tightening torques: 20 Ncm

MIC 6.Ms3.750

**2-CONNECT®-DIM ANALOG®**

KS 62 = 1,6 mm

2CON 5.DIM.160



**SLEEVE** (belongs to DIM ANALOG®)  
plaster model

DIM 5.XSL.A30

**2-CONNECT®-IMPLANT PICK-UP**

integrated screw  
KS 62 = 1,6 mm

2CON TR-00031.6

**2-CONNECT®-LAB ANALOG**

KS 62 = 1,6 mm

2CON 5.La1.600

