





OVERVIEW PLATFORMS:





3,4 mm



4,1 mm



5,0 mm





	Titanium Base (engaged) incl. Abutment Screw	I 800
	Scanbody 3D-Guide® for intra + extra oral scanning	I 9.S3D3.400
1	Lab Analog	I 50
	Implant Pick Up for open-tray with integrated screw	I TR-00013.4
	Abutment Screw recommended tightening torques 35 Ncm	l 61
	DIM-ANALOG®	I 5.DIM.340
	Sleeve (for DIM-ANALOG®) for plaster model	DIM 5.XSL.A10





	Titanium Base (engaged) incl. Abutment Screw	l 810
	Scanbody 3D-Guide® for intra + extra oral scanning	l 9.S3D4.150
1	Lab Analog	l 51
4	Implant Pick Up for open-tray with integrated screw	l TR-00024.1
	Abutment Screw recommended tightening torques 35 Ncm	l 61
	DIM-ANALOG®	I 5.DIM.410
	Sleeve (for DIM-ANALOG®) for plaster model	DIM 5.XSL.A20





Titanium Base (engaged) incl. Abutmentscrew	1820
Scanbody 3D-Guide® for intra + extra oral scanning	I 9.S3D4.150
Lab Analog	I 52
Abutment Screw recommended tightening torques 35 Ncm	I 61
DIM-ANALOG®	I 5.DIM.500
Sleeve (for DIM-ANALOG®) for plaster model	DIM 5.XSL.A30





NOTES



EUROPE

NT-Trading GmbH & Co. KG

G.-Braun-Str. 18 76187 Karlsruhe Germany

E info@nt.dentalW www.nt.dental

NORTH AMERICA

NT-US, Inc.

132 Pepes Farm Road, Unit 3 Milford, CT 06460 USA

E info@nt.dental (general information)E order-us@nt.dental (for orders)

W usa.nt.dental

2024_NT_US_Serienbroschüre_I_EN_REV01

