


BRE-SERIES

compatible to:

Bredent Medical[®]
Sky[®]



PLATFORMS:



3,5 mm



4,0 mm



TITANIUM BASE

incl. screw

BRE 80350000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

BRE 9.S3D3.540



NT-MILL LAB ANALOG

BRE 5.La3.540




IMPLANT PICK-UP

for open-tray with integrated screw

BRE TR-00013.5



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

BRE 603540



DIM ANALOG®

BRE 5.DIM.354













SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A20





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>BRE 10.Mi3.500.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>BRE 9.S3D3.540</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>BRE 5.La3.540</p>	 <p>groove</p>
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>BRE 6.Sc3.540</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>BRE 603540</p>	
<p>DIM ANALOG®</p>	<p>BRE 5.DIM.354</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A20</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.200.000</p>	
<p>I-BUTMENT individual milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

BRE
SERIES



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	BRE-2CON 8035001S
GH 2,5 mm	BRE-2CON 8035002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	BRE-2CON 8035001M
GH 2,5 mm	BRE-2CON 8035002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 20 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw

KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm


2CON 5.La1.800



**2-CONNECT®-ABUTMENT SET (ANGLED)**

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm

BRE-2CON 80350A3S

**2-CONNECT®-ABUTMENT**

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

BRE-2CON 80350A3M

**2-CONNECT®-SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9.S3D1.600

**2-CONNECT®-COPING**

Female Part

2CON F-U

**COPING-SCREW**

Recommended tightening torques: 15 Ncm

KS 62

**MALE PART SCREW**

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

BRE 6.Ms3.540

**2-CONNECT®-DIM ANALOG®**

KS 62 = 1,6 mm

2CON 5.DIM.160

**SLEEVE** (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30

**2-CONNECT®-IMPLANT PICK-UP**

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6

**2-CONNECT®-LAB ANALOG**

KS 62 = 1,6 mm

2CON 5.La1.600





TITANIUM BASE

incl. screw

BRE 80400000



SCAN BODY 3D-GUIDE®

for intra + extra oral scanning

BRE 9.S3D3.540



NT-MILL LAB ANALOG

BRE 5.La3.540




IMPLANT PICK-UP

for open-tray with integrated screw

BRE TR-00013.5



ABUTMENT SCREW



Overview of all screwdrivers on pages 344 and 345

BRE 603540



DIM ANALOG®

BRE 5.DIM.354













SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A20





<p>NT-PREFORM® incl. coloured milling screw</p>	<p>BRE 10.Mi4.000.05</p>	
<p>SCAN BODY 3D-GUIDE® for intra + extra oral scanning</p>	<p>BRE 9.S3D3.540</p>	
<p>NT-MILL LAB ANALOG for NT-Preform®</p>	<p>BRE 5.La3.540</p>	
<p>COLOURED MILLING SCREW for NT-Preform®</p>	<p>BRE 6.Sc3.540</p>	
<p>ABUTMENT SCREW</p>  <p>Overview of all screwdrivers on pages 344 and 345</p>	<p>BRE 603540</p>	
<p>DIM ANALOG®</p>	<p>BRE 5.DIM.354</p>	
<p>SLEEVE (belongs to DIM ANALOG®) plaster model</p>	<p>DIM 5.XSL.A20</p>	
<p>ABUTMENT GRIP for NT-Holder fixture</p>	<p>W 12.200.000</p>	
<p>I-BUTMENT individual milled by NT-iDent</p> <p>Not available in the USA</p>	<p>TZ15.TiO.Si1</p>	

BRE
SERIES



2-CONNECT®-ABUTMENT SET (STRAIGHT)

incl. 2-CONnect®-Coping and Coping-screw

GH 0,5 mm	BRE-2CON 8040001S
GH 2,5 mm	BRE-2CON 8040002S



2-CONNECT®-ABUTMENT

Male Part

GH 0,5 mm	BRE-2CON 8040001M
GH 2,5 mm	BRE-2CON 8040002M



2-CONNECT®-SCAN BODY 3D-GUIDE®

for intra + extra oral scanning
KS 61 = 1,8 mm

2CON 9.S3D1.800



2-CONNECT®-COPING

Female Part

2CON F-U



COPING-SCREW

Recommended tightening torques: 20 Ncm

KS 61



2-CONNECT®-DIM ANALOG®

KS 61 = 1,8 mm

2CON 5.DIM.180



SLEEVE (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30



2-CONNECT®-IMPLANT PICK-UP

integrated screw
KS 61 = 1,8 mm

2CON TR-00021.8



2-CONNECT®-LAB ANALOG

KS 61 = 1,8 mm


2CON 5.La1.800



**2-CONNECT®-ABUTMENT SET (ANGLED)**

incl. 2-CONnect®-Coping, mounting aid,
Coping-screw and Male Part Screw, angled 20°,
Gingiva Height 2,5 mm

BRE-2CON 80400A3S

**2-CONNECT®-ABUTMENT**

angled 20°, Gingiva Height 2,5 mm
Male Part incl. screw,
mounting aid

BRE-2CON 80400A3M

**2-CONNECT®-SCAN BODY 3D-GUIDE®**

for intra + extra oral scanning
KS 62 = 1,6 mm

2CON 9.S3D1.600

**2-CONNECT®-COPING**

Female Part

2CON F-U

**COPING-SCREW**

Recommended tightening torques: 15 Ncm

KS 62

**MALE PART SCREW**

for angled 20° 2-CONnect® Abutment
Recommended tightening torques: 20 Ncm

BRE 6.Ms3.540

**2-CONNECT®-DIM ANALOG®**

KS 62 = 1,6 mm

2CON 5.DIM.160

**SLEEVE** (belongs to DIM ANALOG)

plaster model

DIM 5.XSL.A30

**2-CONNECT®-IMPLANT PICK-UP**

integrated screw
KS 62 = 1,6 mm

2CON TR-00031.6

**2-CONNECT®-LAB ANALOG**

KS 62 = 1,6 mm

2CON 5.La1.600

